

**GOVERNMENT OF ASSAM
OFFICE OF THE DIRECTOR OF MEDICAL EDUCATION, ASSAM
SIXMILE, KHANAPARA, GUWAHATI-22**

No.DME/396/2010/1412,

Dated: 13/01/2011

CORRIGENDUM

Please read KV selection '90 to 140 KV' as from 90 to 140KV.

Civil works for the 16-slice CT Room (to be done by the vendor)

1. Construction to be reinforced as per drawing.
2. The floor must be vitrified tiles.
3. The paint of wall must be plastic paint.
4. Ceiling type must be false ceiling with semi perforated boards and Aluminum T. Channel. (in Gantry, Console, workstation and Doctors room)
5. **Air conditioner :**
Total Requirements – 6 Nos
Gantry 1.5 tone - 3 Nos - (The Gantry area should cooled in the range between 18°C - 20° C).
Console 1 tone - 3 Nos - (1 in Console + 1 in Doctors Room + 1 in patients area, cooling 18°C - 20° C needed).
6. Earthing to be done by vendor/ Chemical earthing is preferred
7. Light fixture should be mentioned for proper illumination.
8. Fire alarm and portable fire extinguisher to be fixed in the machine and control room.
9. Extra generator room to be build with a extra entry door for the generator room.
10. Furniture requirements :
 - A. Computer Chair - 2 Nos. (revolving)
 - B. Chair - 10 (fixed)
 - C. Table L-Type - 1 Nos.
 - D. Standard Table - 3 Nos. (3" X 2")
 - All software are needed in console, workstation, for faster completion of patients screening and MPR etc.
 - Generator for whole system
 - UPS for console, workstation & pressure injector
 - Stabilizer for A.C.
11. Penalty: up time up to 95 % or better no penalty, for 1 % down time vendor will have to extend warranty by one week.
12. The work of installation and handing over must be completed within a period of 60 (sixty) days from the opening of LC.

SD/-PROF. D.HAZARIKA.
Director of Medical Education, Assam

Memo No.DME/396/2010/1413-17,

Dated: 13/01/2010

Copy to-

1. The Commissioner & Secretary to the Govt. of Assam, Health & F.W. (B) Deptt., Dispur, Guwahati-6 for information.
2. The Mission Director, NRHM, Assam. He is requested kindly to upload the Corrigendum Notice in the NRHM, Assam's website. Soft copy enclosed.
3. The State Informatics Officer, Block F, Assam Secretariat, Dispur, Guwahati-06. He is requested to upload the Corrigendum Notice in the Assam Govt. Website. Soft copy enclosed.
4. The P.S. to the Hon'ble Minister, Health & F.W., Assam for kind appraisal of the Hon'ble Minister.
5. The Notice Board of this Directorate.

Director of Medical Education, Assam