



OFFICE OF THE MISSION DIRECTOR
NATIONAL HEALTH MISSION, ASSAM

SAIKIA COMMERCIAL COMPLEX, CHRISTAN BASTI, G.S ROAD, GUWAHATI-781005

PH.NO : 0361-2363062 ; TELE FAX : 0361-2363058

Website: www.nrhmassam.in

e-mail: misnrhm.assam@gmail.com

File No: NHM/PROC/ Dental-Chair/2988/2016-17/279

Date: 7/4/17

CORRIGENDUM No.1

e-TENDER FOR ENTERING INTO RATE CONTRACT FOR DENTAL CHAIRS AND ACCESSORIES

This has reference to the e-tender No: NHM/PROC/ Dental-Chair/2988/2016-17/22067 Date: 16/03/17 for entering into rate contract for “Supply of Dental Chairs and Accessories” to health institutions in different districts of Assam. The following amendment may be taken note of prior to submission of Bids:

1. Clause 3 (Point No:3) “ELIGIBILITY CRITERIA” is amended to read as follows:

3 ELIGIBILITY CRITERIA:

C. Manufacturer participating as bidder or manufacturer from whom the Hospital Equipments/Instruments would be sourced by the dealer should have an average Annual Turnover of **Rs.6.00 Crore (Rupees Six Crore only)** in the past 3 financial years 2013-14, 2014-15 & 2015-16.

2. Clause 5 of Techno-COMMERCIAL BID - Cover “A” is amended to read as follows:

I. USFDA/ CE, ISO 9001:2008 & ISO 13485:2003 Certificates

- i) Attested copies of valid **USFDA/CE** (only of manufacturers whose products have been offered) for Dental Chairs, Air Compressors and Servo Voltage Stabilizers.
- ii) Manufacturers whose items have been quoted should have valid **IEC 60601/IS-13450** as electrical safety standards for electrical items.
- iii) Attested copies of valid **ISO 9001:2008 and ISO 13485:2003** certificates (of manufacturers whose products have been offered) for Dental Chairs.
- iv) Attested copies of valid **ISO 9001:2008** certificates (of manufacturers whose products have been offered) for the Accessories.

J. Annual Turnover

Annual turnover statement from sales of Hospital Equipments/Instruments for past 3 years i.e. 2013-14,2014-15 and 2015-16 in the format given in **Annexure-III A and Annexure-III B** certified by the Auditor/Chartered Accountant. The certificate of Auditor/Chartered Accountant must be supported by attested copies of “Profit & Loss Statement” and “Balance sheet” of past three financial years. It is to be noted that such statement of both the bidder and the manufacturer (in case the bidder is not a manufacturer) must be submitted.

M. Undertaking for providing of logo of Bidder:

Undertaking (as per **Annexure-VI**) for embossment of NHM Assam logo on items like Dental Chairs, Compressors and Voltage Stabilizer. For other small accessories like forceps, Needle, files etc., bidder should print or emboss or affix the Logo in the outer cover of the unit packets.

3. Clause 20: of the tender document is amended to read as follows:

LOGOGRAMS:

- A. Tenders for the supply of Dental Chairs and Accessories, shall be considered only if the Bidder gives undertaking in his tender that the items will be supplied with the logogram either printed or embossed or affixed as specified at Clause 5 (Point M).
- B. Failure to supply Dental Chairs and Accessories, without the logogram will be treated as breach of the terms of agreement.

4. Annexure XII “Technical cum Compliance Specification Sheet” is amended to read as follows:

ANNEXURE XII (Revised)

Technical cum Compliance Specification Sheet: (TO BE SUBMITTED BY THE MANUFACTURER)

1. Dental Chair with Accessories

SI	Name	Technical Specs quoted by bidder	Deviation if any
1 Description of Function			
1.1	Dental Chair is required for dental examination and dental procedures.		
2 Operational Requirements			
2.1	Physiological Dental Chair (Basic) operated by electricity with motorized head rest, up and down movement for		

SI	Name	Technical Specs quoted by bidder	Deviation if any
	perfect adjustment of head position.		
3 Technical Specifications			
3.1	<p><u>Dental Chair:</u> Dental unit with over the patient modular delivery system</p> <ul style="list-style-type: none"> • It should have two 3 way syringes (tip autoclavable) • It should have high speed Air Rotor terminal supplied with original CE mark 380000 rpm hand piece and having water control on coupling supplied with the hand piece. (Hand piece make: - NSK/Kavo/W&H/Bien Air). • Should have one brush (extra 2 set of carbon brush)/ brushless micro motor (40000rpm±10%) terminal with straight and contra angle hand pieces (Hand Piece make: NSK/Kavo/W&H/Bien Air) • It should have one in-built Piezon Ultrasonic Scaler (frequency 28-36 KHz) with 8 scaler tips. • It should have LED Light with sensor (In the range of 30,000-40,000 LUX) • It should High and low motorized suction with auto drain and flush system. • Removable auxiliary tray (stainless steel) • Auto connection for tumbler and spitoon • It should have LED based X-ray viewer. • It should be provided with one doctor's stool with adjustable backrest tilt including an adjustable for foot rest. 		
4 System Configuration Accessories, spares and consumables			
4.1	<ul style="list-style-type: none"> • System as specified • All consumables required for installation and standardization of system to be given free of cost. <p><u>Accessories:</u></p> <p>1. Compressor –1 no. with the following specification.</p> <ul style="list-style-type: none"> • Air Compressor Oil Free • Voltage (V) : 90-280 • Max.Air Flow (L/Min) : 152 • Air Flow at Bar : 50 • Max.Pressure (Bar) : 8 • Tank (Litre) : 30 		

SI	Name	Technical Specs quoted by bidder	Deviation if any
	<ul style="list-style-type: none"> • Noise level (dB) : <70 • Dimension (mm) : 410 x 410 x 650 • Weight (kg) : 33 <p>2. Kidney Tray 10", Tweezer , Explorer, Mouth Mirror Top & Handle-2 sets (ISO 9001:2008 certified)</p> <ol style="list-style-type: none"> 1. Extraction Forceps (full set)-Permanent dentition Set of 14 & Primary dentition Set of 7-2 sets (ISO 9001:2008 certified) 2. Elevators Set of 9 (Cross Bur, Cryer's, Straight, Coupland)-2 sets (ISO 9001:2008 certified) 3. Periosteal elevators (Small & Medium)-1 set (ISO 9001:2008 certified) 4. B.P handle-No:3 and 4 - 1 set (ISO 9001:2008 certified) 5. Needle Holder -1 No. (ISO 9001:2008 certified) 6. Suture-1 No. 7. Needle (Size 17,18,19,20)-1 Set 8. Restorative Instrument Set of 13 with Amalgam Carrier-1 set (ISO 9001:2008 certified) 9. Burs for cavity preparation Set of 12 (Round, Inverted, Tapered Fissure, Flat Fissure) with bur box-1 Set (ISO 9001:2008 certified) 10. Cotton holder-1 no (ISO 9001:2008 certified) 11. Mortar, Pestle & Amalgam Squeezer-1 set (ISO 9001:2008 certified) 12. Dappen dish-1 no (ISO 9001:2008 certified) 13. Tofflemire retainer with Band-1 Set (ISO 9001:2008 certified) 14. Wedges Pkt of 50 (Plastic & Wooden)-1 Pkt (ISO 9001:2008 certified) 15. Endodontic files (ISO 9001:2008 certified)-1 packet comprising of: <ul style="list-style-type: none"> • K File No 10 21 mm & 25mm- 1 Set • K File No 15-40 21 mm & 25mm- 1 Set • K File No 45-80 21 mm & 25mm- 1 Set • H File No 10 21 mm & 25mm- 1 Set • H File No 15-40 21 mm & 25mm- 1 Set • H File No 45-80 21 mm & 25mm- 1 Set • Reamer 15-40 21 mm & 25mm- 1 Set • Reamer 45-80 21 mm & 25mm- 1 Set • Plugger 15-40 21 mm & 25mm- 1 Set • Spreader 15-40 21 mm & 25mm- 1 Set 		

SI	Name	Technical Specs quoted by bidder	Deviation if any
	<p>16. Hand scalers double end set of 8 -1 set (ISO 9001:2008 certified)</p> <p>17. Mixing Bowl (Small & Big)-1 Set. (ISO 9001:2008 certified)</p> <p>18. Mixing spatula-for plaster & Alginate (Straight & Curve)-1 Set (ISO 9001:2008 certified)</p> <p>19. Impression Trays(full set)-Perforated (Dentulus & Edentulus)-1 Set (ISO 9001:2008 Certified)</p> <p>20. Wax knife, Wax spatula, Hot Plate-1 set (ISO 9001:2008 Certified)</p> <p>21. Sprit Lamp-1 No (ISO 9001:2008 certified)</p>		
5 Environmental factors			
5.1	<ul style="list-style-type: none"> • The unit shall be capable of operating continuously in ambient temperature of 10-40deg C and relative humidity of 15-90% • Complete installation of the system including water input and drainage system has to be installed 		
6 Power Supply			
6.1	<ul style="list-style-type: none"> • Power input to be 90-280VAC, 50Hz • 3 KVA Servo Voltage stabilizer of appropriate ratings meeting ISI Specifications.(Input 90-280 V and output 220-240 V and 50 Hz) 		
7 Standards, Safety and Training			
	<p>Certifications</p> <ul style="list-style-type: none"> • Should be USFDA/ CE approved product • Manufacturer/Supplier should have ISO 9001:2008 and ISO 13485:2003 certification for quality standards. • Electrical safety conforms to standards for electrical safety IEC-60601-1-2 / IS-13450 for the electrical items. <p>Warranty</p> <ul style="list-style-type: none"> • 3 years for Dental Chair including Hand-piece, Compressors and Servo Voltage Stabilizer. • 1 year for other accessories <p>Quarterly visit to site, within warranty period including training of user on maintenance</p>		
8 Documentation			
	<ul style="list-style-type: none"> • User/Technical/Maintenance manuals to be supplied in English. • Certificate of calibration and inspection. • List of important spare parts, hand pieces, and 		

SI	Name	Technical Specs quoted by bidder	Deviation if any
	accessories with their part number and costing <ul style="list-style-type: none"> • Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. 		

All other terms and conditions of the tender referred to above shall remain unchanged. The start date and end date of bid submission is extended to **18/04/2017 from 10:00 AM & 24/04/2017 up to 2:00 PM** respectively.

**Sd/-
Mission Director, NHM, Assam**